Contract Number: 710419

Certificate: SGSNA/18/SUW/00013

Test Report Number: 624870/01

Certified Product: Earigator - Cerumen Management System

Trademark (if any):

Model(s):

NT0001-1, NT0001-2

Technical Data: Type B applied part, Class I;

NT0001-1: 100-120Vac, 50/60Hz, 10A; NT0001-2: 220-240Vac, 50/60Hz, 5A

Name and Address of Applicant:

Nupur Technologies, LLC

1576 Sweet Home Road 14228, Buffalo, New York,

United States of America

Name and Address of Manufacturer:

Nupur Technologies, LLC

1576 Sweet Home Road 14228, Buffalo, New York,

United States of America

Name and Address of Factory:

Nupur Technologies, LLC

1576 Sweet Home Road 14228, Buffalo, New York,

United States of America

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com. The product is certified according to ISO/IEC Guide 17067, Conformity assessment - Fundamentals of product certification, System 3, and in accordance with

ANSI/AAMI ES60601-1:2005 +C1 (2009) +A1 (2012) +A2 (2010) CAN/CSA C22.2 No. 60601-1:14

Authorized by

Effective date: 22 January 2018

aul Krauss Certifier

Consumer and Retail Services, a division of SGS North America Inc. 620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA t +1 770 570 1800 f +1 770 277 1240 www.sgs.com Page 1 of 1





This certificate is issued by the company under its General Conditions for Certification Services accessible at /www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this Certificate. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.



